

Hydraulic properties of the Hemopump HP31: a study of the downstream pressure distribution

M. Grigioni*, A. Carotti**, C. Daniele*, U. Morbiducci*, C. Del Gaudio*, G. D'Avenio*,
V. Barbaro*, R. Di Donato**

*Laboratory of Biomedical Engineering, Istituto Superiore di Sanità, Viale Regina Elena 299, 00161 Roma, Italy

** "Bambino Gesù" Pediatric Hospital, Roma, Italy

Abstract – The Hemopump was commercialized as an useful tool for the left ventricle assistance. Bioengineers and clinicians showed great interest to develop applications and to analyze its hydraulic behaviour; in this work an application for axial pump in different conditions is presented. A study of the spatial pressure distribution generated by the impeller of the Hemopump is investigated in highly accurate steady-flow conditions.

The experimental set up adopted for this study consists in a plexiglass test pipe (simulating an aortic conduit of 22-mm diameter) and allows the sampling of the pressure at the outlet of the pump in 16 points spaced $\frac{1}{2}$ diameter each other. Keeping fixed the constant head at the inlet of the Hemopump and varying the constant head at its outlet, i. e. afterloads, in 11 step levels, it was possible to draw the characteristic flow curves versus delivered pressure for all the seven speed levels. A pressure range of about 35÷130 mmHg and a flow range of about $-0.7 \div 3.7$ l/min was experimented.

The results show that the flow delivered by the Hemopump is fully developed after 5 pressure taps (about 55 mm), with no further varying along the test chamber. These data could be used to optimize the setting up of clinical experimental procedures.

Keywords – Hemopump, spatial pressure distribution

I. INTRODUCTION

In recent years, the need has grown for mechanical circulatory support of the failing heart. Micro-axial blood pumps have been revealed as an useful tool either for short term or for chronic left ventricular assistance [1-7].

The Hemopump (Medtronic, Minneapolis, USA) is a device based on the principle of the screw pump: it furnishes, by conversion from electric energy, rotating energy to a high speed rotor [8]. The elevated speed rotation accelerates blood transferring it from the low pressure inlet to the high pressure outlet. The Hemopump is a device for left ventricular assistance having an axial pump, with an aspiration silicone cannula, connected by a flexible drive cable to an electromagnetic motor, this one being controlled by the Hemopump drive unit [7, 8]. The drive unit provides seven different speed rotation levels, trying to maintain at a constant value the pump rotation speed for each available setting controlling the motor current erogation in consequence of the changing pressure and flow working conditions. This current-controlled method does not allow to desume the pump flow from the selected rotation speed, with the consequence that the level of perfusion could be unpredictable, depending on preload and afterload conditions.

Our group used this axial pump to realize an extracorporeal circuit in the case of ewe's fetal cardiosurgery trials [9].

This device, in its normal use, is pushed through the aorta in the left ventricle and fixed in transvalvular position. Object of the present study is an in vitro test of the hydraulic performances of the HP31 in a cylindrical test chamber, to investigate the pressure distribution at different distances from the edge of the pump impeller. Fixing the constant head at the inlet of the Hemopump and varying the constant head at its outlet, it was possible to draw the characteristic flow curves versus delivered pressure for all the available pump rotating speed levels, to be used as a control grid in defining experimental procedures (cannulation, circuit length, etc.) such as in [9,10].

II. MATERIALS & METHODS

The experimental set up adopted for the investigation of the pressure distribution distal to the axial pump outlet is shown in Fig. 1. The hydraulic circuit is equipped with a plexiglas test pipe (22 mm diameter) with 16 equispaced taps (1/2 diameter length) for the pressure measurements. The latter were performed using water columns for each pressure tap (see Fig. 1) at the resolution of 0.05 cmH₂O. The Hemopump, mounting the aspiration cannula, was inserted co-axially within the test chamber, while two reservoirs were connected by plastic tubes.




Fig 1. Experimental set up

Report Documentation Page

Report Date 25 Oct 2001	Report Type N/A	Dates Covered (from... to) -
Title and Subtitle Hydraulic Properties of the Hemopump HP31: A Study of the Downstream Pressure Distribution		Contract Number
		Grant Number
		Program Element Number
Author(s)		Project Number
		Task Number
		Work Unit Number
Performing Organization Name(s) and Address(es) Laboratory of Biomedical Engineering Istituto Superiore di Sanita Viale Regina Elena 299 00161 Roma, Italy		Performing Organization Report Number
Sponsoring/Monitoring Agency Name(s) and Address(es) US Army Research, Development & Standardization Group (UK) PSC 802 Box 15 FPO AE 09499-1500		Sponsor/Monitor's Acronym(s)
		Sponsor/Monitor's Report Number(s)
Distribution/Availability Statement Approved for public release, distribution unlimited		
Supplementary Notes Papers from 23rd Annual International Conference of the IEEE Engineering in Medicine and Biology Society, Oct 25-28, 2001, held in Istanbul, Turkey. See also ADM001351 for entire conference on cd-rom, The original document contains color images.		
Abstract		
Subject Terms		
Report Classification unclassified		Classification of this page unclassified
Classification of Abstract unclassified		Limitation of Abstract UU
Number of Pages 3		

The hydrodynamic behaviour of the pump was investigated in an in vitro model filled with a 33% solution of water and glycerol, simulating blood viscosity.

The study was performed with respect to a feasible range of pressures delivered at pump's outlet [11, 12] using gravimetric heads, ranging from 35 to 130 mmHg; the test was performed investigating all the seven pump speed levels.

A centrifugal pump (Iwaki Co. Ltd) between reservoir 2 (output constant head) and 1 (input constant head) has been used, to supply constant head tanks.

Thus the hydraulic behaviour of the axial pump was investigated fixing the piezometric height of the reservoir 1 (input constant head), keeping at a constant piezometric height the reservoir 2 (output constant head), for all the seven pump speed levels; this configuration was repeated for eleven different piezometric heights of the reservoir 2, in elevation steps of 12.5 cm of water-glycerol solution (approximately 10 mmHg).

The above-described set-up allowed us to measure the difference between the pressure at the outlet of the pump (pressure value measured at the tap farthest from the pump's outlet) and the pressure proximal to the inlet of the aspiration cannula (calculated assuming a poiseuillian fluid motion in the adduction tube).

Pressure measurements have been done with a gravimetric system of very high precision, so that the standard deviations resulted of the same order of accuracy of the measurement system itself.

Flow measurements were performed with an electromagnetic flowmeter (Endress+Hauser), having 0.1% of accuracy in detecting fluid velocity.

III. RESULTS

Figures. 2, 3, 4 show the pressure distributions along the test section (16 pressure taps) for three different speed levels (speed level 2, 4 and 7 respectively), for all the eleven investigated heights of the output constant head.




Fig 2. Pressure distributions along the test section at speed level 2 (from \blacklozenge =10 mmHg to \blacksquare =110 mmHg, step=10 mmHg)

It can be noticed how, for all the speed levels, the measured pressures showed an increasing trend as for the first 5, 6 pressure taps distal to the edge of the pump impeller. It is possible to underline how a difference in the piezometric height of the reservoir 2 (output constant head) sensibly modify the pressure delivered by the axial pump. This could

be relevant when designing the length of an external circuit as well as of the cannulation.

Figure 5 shows the flow rate versus delivered pressure curves for all the eleven piezometric heights of the reservoir 2; it is possible to observe how, at all pump speed levels, a reduction of the afterload was associated to an increasing antegrade pump flow.




Fig 3. Pressure distributions along the test section at speed level 4 (from \blacklozenge =10 mmHg to \blacksquare =110 mmHg, step=10 mmHg)

Figure 5 also shows how retrograde flow was obtained at higher afterloads and lower speed levels (the three lower ones).

In Figure 6 the relationship is shown between the flow rates and the differences between the pressure at the outlet of the test section (measured at the tap farthest from the pump's outlet) and the pressure proximal to the inlet of the aspiration cannula.




Fig 4. Pressure distributions along the test section at speed level 7 (from \blacklozenge =10 mmHg to \blacksquare =110 mmHg, step=10 mmHg)

IV. DISCUSSION

In this paper we measured the pressure spatial distribution distally to the edge of an HP31 impeller [6], in a cylindrical tube simulating an arterial site. This allowed us to quantify, under specific controlled conditions, the transitional length of the fluid motion, i. e., at which distance far from the edge of the axial pump impeller the flow can be considered fully developed; at the same time, pressure measurements can be considered stabilized.

We observed how, for all the investigated piezometric heights of the reservoir connected to the line of return of the experimental hydraulic circuit (output constant head), at a distance of about 55 mm away, distally to the edge of the

Hemopump impeller (i. e., the first five pressure taps, equivalent to a distance of about three vessel diameters of the cylindrical test chamber), pressure reached an almost constant value, no further varying along the test chamber.




Fig 5. Flow rate versus delivered pressure curves relative to the considered eleven piezometric heights of the reservoir 2.




Fig 6. Flow rate versus pressure differences curves for all the eleven piezometric heights of the reservoir 2.

We observed how the axial pump is sensible to the afterload value: according to [11, 12], pump performance was significantly improved by afterload reduction. This means that in a beating heart, during the mechanical assistance, the blood flow provided by the pump is modulated by the arterial loading conditions, together with the ventricular contribution.. As a consequence, the current-controlled method governing the HP does not allow to desume the pump flow from the selected rotation speed, with the consequence that the level of perfusion remains not exactly predictable.

The level of indetermination in the flow rate delivered by the axial pump is an in vivo undesirable feature, so that in vitro hydraulic characterizations of the HP31 could furnish useful estimations for clinical set up.

The pressure field at pump's outlet, for each afterload level and for each speed level, was investigated as a measure of the Hemopump's capability to convert kinetic energy into pressure; we demonstrated that this capability depends on the hydraulic characteristics of the circuit in which the pump is investigated. The data suggest that in vitro behaviours can be used to accurately set the experimental assistance and to correctly evaluate the results as in [9, 10].

The results presented in this paper also showed how for high afterload and low speed levels the fluid flow can even become

negative, indicating a retrograde flow through the axial pump. We measured retrograde pump flow in correspondence of the first three pump speed levels, suggesting that care must be paid to avoid this occurrence in vivo, in order to maintain the efficiency of the mechanical assistance.

REFERENCES

- [1] B. P. Meyns, P. T. Sergeant W. J. Daenen and W. J. Flameng, "Left ventricular assistance with the transthoracic 24F Hemopump for recovery of the failing heart" *Ann. Thorac. Surg.* 1995, vol. 60, pp. 392-397.
- [2] R. W. Smalling, D. B. Cassidy, R. Barrett, B. Lachterman, P. Felli and J. Amirian "Improved regional miocardial blood flow, left ventricular unloading, and infarct salvage using an axial-flow, transvalvular left ventricular assist device" *Circulation* 1992, vol. 85, pp 1152-1159.
- [3] M. E. Merhige, R. W. Smalling, D. B. Cassidy, R. Barrett, G. Wise, J. Short and R. K. Wampler "Effect of the hemopump left ventricular assist device on regional myocardial perfusion and function: reduction of ischemia during coronary occlusion" *Circulation* 1989, vol. 80 (suppl III), pp III-158-III-166.
- [4] P. F. Wouters, S. Sukehiro, T. Mollhoff, et al "Left ventricular assistance using a catheter-mounted coaxial flow pump (Hemopump) in a canine model of reginal myocardial ischaemia" *Eur. Heart J.* 1993, vol. 14, pp. 567-575
- [5] R. K. Wampler, O. H. Frazier, A. M. Lansing, R. W. Smalling, J. M. Niklas, S. J. Phillips, R. A. Guyton and L. A. R. Golding "Treatment of cardiogenic shock with the Hemopump left ventricular assist device" *Ann. Thorac. Surg.* 1991, vol. 52, pp. 506-513.
- [6] G. D. Dreyfus "Hemopump 31, the sternotomy Hemopump: clinical experience" *Ann. Thorac. Surg.* 1996, vol. 61, pp. 323-328.
- [7] R. K. Wampler, B. A. Baker and W. M. Wright "Circulatory support of cardiac interventional procedures with the Hemopump™ cardiac assist system" *Cardiology* 1994, vol. 84, pp. 194-201.
- [8] K. C. Butler, J. C. Moise and R. K. Wampler "The Hemopump® - a new cardiac prothesis device" *IEEE Transactions on Biomedical Engineering*, vol. 3, no. 2, pp.194-196, February 1190.
- [9] Grigioni M, Carotti A, Daniele C, D'Avenio G, Morbiducci U, Iannace E, Albanese S B, Costa D, Formigari R, Ferretti E, Di Donato R M. "Extracorporeal circulation in ewe's foetus: towards a reliable foetal cardiac surgery protocol". *Int J Artif Organs* 2000; 23: 189-198.
- [10] Reddy VM, Liddicoat JR, Klein JR, McElhinney DB, Wampler RK, Hanley FL. "Fetal cardiac bypass using an in-line axial flow pump to minimize extracorporeal surface and avoid priming volume" *Ann Thorac Surg* 1996;62:393-400.
- [11] T. Sieß, H. Reul and G. Rau "Concept, realization and first in vitro testing of an intraarterial microaxial blood pump" *Artificial Organs* 1995, vol. 19, no. 7, pp.644-652.
- [12] M. Lachat, C. Jaggy, B. Leskosek, L. von Segesser, G. Zund, P. Vogt and M. Turina. "Hemodynamic properties of the Hemopump HP14 ". *Int J Artif Organs* 1999; 22: 155-159.